



EU-Declaration of Conformity for Medical Device Class IIb

Hamburg, 2023-11-07

Object of the declaration: Mikrobac Virucidal Tissues

Mikrobac Virucidal Tissues		
Pack size	Article number BODE	Article number HARTMANN
Flowpack 80 Tissues	981531	981531

We herewith declare under our sole responsibility that the medical devices listed above, first placed on the market by BODE Chemie GmbH, comply with the applicable provisions, in particular, the

- General Safety and Performance Requirements of Regulation (EU) 2017/745 of the European Parliament and of the Council of 5. April 2017 on medical devices.

The objects of the declaration have been identified as medical devices in risk class IIb according to classification rule 1 and rule 16 in Annex VIII of Regulation (EU) 2017/745.

The conformity assessment procedure according to Article 52 (4) and Annex IX has been performed and the Technical Documentation is kept available.

The conformity assessment procedure is under the supervision of the Notified Body:

MEDCERT Zertifizierungs- und Prüfungsgesellschaft für die Medizin GmbH
Pilatuspool 2
20355 Hamburg
Germany
Identification No. 0482
Certificate No. 0523GB448210329A


(High-Level) Intended Purpose:
Disinfection of non-invasive and invasive medical devices

Basic UDI-DI: 40316783779ML
Single Registration Number: DE-MF-000005851

BODE Chemie GmbH



Britta Glogau
Director Research & Development



Raphael Bohner
Head of Quality

Valid until: 2025-11-07

